

TEST REPORT EN 12183:2014 TUV SUD Test report for manual wheelchairs - Requirements and test methods	
Report reference No.	713163847
Date of issue	2019-11-11
Project handler.....	Jens Meyer
Testing laboratory	TÜV SÜD Product Service GmbH
Address	Masurenweg 1-3, D-30163 Hanover
Testing location.....	as above
Client	Invacare International GmbH
Client number.....	5010295184
Address	Benkenstrasse 260, CH-4108 Witterswil
Contact person.....	Mrs. Mareike Lehr
Standard.....	This TUV SUD test report form is based on the following requirements: EN 12183:2014
TRF originated by	TUV SUD Product Service GmbH, Mr. Jens Meyer (product specialist)
Copyright blank test report	This test report is based on the content of the standard (see above). The test report considered selected clauses of the a.m. standard(s) and experience gained with product testing. It was prepared by TUV SUD Product Service GmbH. TUV SUD Group takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.
Scheme	<input type="checkbox"/> TÜV Mark, <input checked="" type="checkbox"/> without certification
Non-standard test method.....	<input checked="" type="checkbox"/> No <input type="checkbox"/> Yes, see details under Summary
National deviations	---
Number of pages (Report).....	53
Number of pages (Attachments)	---
Compiled by	Approved by.....
(+ signature)	(+ signature)
 SIGN-ID 300456	 SIGN-ID 300458

Test sample.....	HAN-447771-1, HAN-448135-1, HAN-451403-1	
Type of test object.....	Manual wheelchair	
Trademark.....	Küschall <small>UNLIMIT YOUR WORLD</small>	
Model and/or type reference	Küschall K-Serie	
Rating(s).....	---	
Manufacturer	Invacare France Operations	
Manufacturer number.....	1300032457	
Address	Route de Saint Roch, FR-37230 Fondettes	
Sub-contractors/ tests (clause).....	---	
Name.....	---	
Order description.....	<input checked="" type="checkbox"/>	Complete test according to TRF
	<input type="checkbox"/>	Partial test according to manufacturer's specifications
	<input type="checkbox"/>	Preliminary test
	<input type="checkbox"/>	Spot check
	<input type="checkbox"/>	
Date of order.....	2019-07-02	
Date of receipt of test item.....	2019-09-16, 2019-09-17, 2019-10-02 (test samples) 2019-10-14, 2019-10-23 (technical documentation)	
Date(s) of performance of test.....	2019-09-23 to 2019-11-11	
Test item particulars:	---	
Attachments:	---	
General remarks:	<p>"(see remark #)" refers to a remark appended to the report. "(see appended table)" refers to a table appended to the report. Throughout this report a comma is used as the decimal separator. The test results presented in this report relate only to the object tested. This report shall not be reproduced except in full without the written approval of the testing laboratory.</p>	

Summary of testing:

Additional information on Non-standard test method(s)

Sub cl..... :---

Page..... :---


Rational..... :---

deviation(s) found

no deviations found

Copy of marking plate:

Example of the type labels



Picture of the product

